

## Service & maintenance checklist

### Door adjustment Water Treadmill

1. Check sealings of doors on any leakage when water rised up in the water treadmill and doors are closed
2. Loose conter nut at door fixation left and right side
3. Slightly turn clockwise nut on any leakage. Rubber of sealing should be even and not pressed complete. Take care that both sides pressed equal against the container and **ONLY** tidened until there is no water leaking through the sealing
4. Tiden the conter nut for propper fixation of the door fixation

#### **Important notice!**

- Any malfunction is to be immediately reported to local service team. Any leakage at pipework, filters, tanks or pumps is to be reported immediately to local service team.